

 Always keep smartphone plugged into charger.

Getting Started

01. Remove the smartphone and charger from box. The phone is sufficiently charged for the first use. Take out Pulse Oximeter (batteries are already placed in).

02. Start your phone, power ON button is on the right side. Hold button until logo appears. Be sure the phone is turned on before you take a reading. (see *diagram 01*)

03. The smartphone and pulse oximeter have already been connected for you. The phone needs to remain on for continued connection.



diagram 01

04. Power on pulse oximeter by pushing the circular button on top. The meter will light up. (see *diagram 02*)




diagram 02

This is not an emergency response unit. If you are experiencing any unusual symptoms, please seek medical help immediately, or dial 911.

 Always keep smartphone plugged into charger.

Getting Started (cont'd)

05. Place your middle, ring or index finger of your left hand into the rubber opening of the oximeter with nail side down. (see *diagram 02 & 03*).




diagram 03

06. After a few seconds, your SpO2 reading will appear on the oximeter display screen. Remove your finger after a few seconds. It will then automatically sync with your smartphone when you hear a sound alert from your smartphone.

Important

Make sure the phone is ALWAYS on charge. Removing the phone from the charger can lead to connection issues.

For additional assistance please visit www.ehiehr.com/rpmpatients

If any issues occur, please contact RPM support at 973-421-5222.

This is not an emergency response unit. If you are experiencing any unusual symptoms, please seek medical help immediately, or dial 911.